


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 UKAS CALIBRATION 8239 Accredited to ISO/IEC 17025:2017	PASS (Portable Appliance Safety Services) Ltd Issue No: 015 Issue date: 18 December 2024	
	1 Wilson Street Thornaby Stockton-On-Tees TS17 7AR United Kingdom	Contact: Mr Ibrahim Ibrahim Tel: +44 (0) 1642 626148 Fax: +44 (0) 870 143 1869 E-Mail: ibrahim@calibrate.co.uk Website: www.calibrate.co.uk

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 1 Wilson Street Thornaby Stockton-On-Tees TS17 7AR United Kingdom	Temperature, Electrical and Pressure	A
Address Parkburn Court Burnbank Hamilton Scotland ML3 0QQ	Dimensional and Electrical	B

Site activities performed away from the locations listed above:

Location details	Activity	Location code
At customer's premises The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Dimensional	C



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL			All electrical calibrations are performed as a comparison against a reference standard	A
DC RESISTANCE				
Fixed value sources for the calibration of measuring instruments	0.1 Ω 0.2 Ω 0.3 Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω 1 G Ω	5.8 m Ω 41 m Ω 41 m Ω 6.0 m Ω 8.2 m Ω 490 $\mu\Omega$ 4.8 m Ω 49 m Ω 980 m Ω 33 Ω 3.1 k Ω 210 k Ω 12 M Ω		
For generating a stimulus that can be applied to measuring instruments also for measuring a stimulus provided by the device being calibrated	0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	17 $\mu\Omega/\Omega + 59 \mu\Omega$ 14 $\mu\Omega/\Omega + 590 \mu\Omega$ 12 $\mu\Omega/\Omega + 740 \mu\Omega$ 12 $\mu\Omega/\Omega + 6.6 m\Omega$ 12 $\mu\Omega/\Omega + 76 m\Omega$ 17 $\mu\Omega/\Omega + 3.3 \Omega$ 58 $\mu\Omega/\Omega + 130 \Omega$ 580 $\mu\Omega/\Omega + 2.8 k\Omega$ 0.58 % + 94 k Ω		
DC VOLTAGE				
Values can be generated for the calibration of measuring instruments	0 mV to 202 mV 202 mV to 1 V 1 V to 2.02 V 2.02 V to 10 V 10 V to 20.2 V 20.0 V to 100 V 100 V to 202 V 202 V to 1020 V	17 $\mu V/V + 2.5 \mu V$ 10 $\mu V/V + 3.6 \mu V$ 10 $\mu V/V + 7.6 \mu V$ 9.8 $\mu V/V + 43 \mu V$ 9.8 $\mu V/V + 72 \mu V$ 14 $\mu V/V + 430 \mu V$ 14 $\mu V/V + 720 \mu V$ 14 $\mu V/V + 2.8 mV$		
For measurement of instrument Outputs	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	5.8 $\mu V/V + 390 nV$ 4.6 $\mu V/V + 430 nV$ 4.6 $\mu V/V + 1.2 \mu V$ 6.9 $\mu V/V + 54 \mu V$ 15 $\mu V/V + 1.3 mV$		



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
DC CURRENT				A
Values can be generated for the calibration of measuring instruments	0 μ A to 202 μ A 202 μ A to 1 mA 1 mA to 2.02 mA 2.02m A to 10 mA 10 mA to 20.2 mA 20.2 mA to 100 mA 100 mA to 202 mA 202 mA to 1 A 1 A to 2.02 A 2.02 A to 10 A 10 A to 20.2 A 20.2 A to 30 A	120 μ A/A + 12 nA 58 μ A/A + 35 nA 58 μ A/A + 49 nA 58 μ A/A + 230 n 58 μ A/A + 440 nA 58 μ A/A + 2.3 μ A 58 μ A/A + 9.0 μ A 150 μ A/A + 36 μ A 150 μ A/A + 100 μ A 350 μ A/A + 590 μ A 350 μ A/A + 760 μ A 580 μ A/A + 4.4 mA		
For measurement of instrument outputs	20 A to 1500 A	0.26 % + 13 mA	Simulation with coil	
	0 μ A to 1 μ A 1 μ A to 10 μ A 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 3 A 3 A to 5 A 5 A to 10 A	24 μ A/A + 48 pA 23 μ A/A + 130 pA 23 μ A/A + 950 pA 23 μ A/A + 6.0 nA 23 μ A/A + 60 nA 40 μ A/A + 630 nA 130 μ A/A + 13 μ A 0.23 % + 750 μ A 0.14 % + 2.6 mA 0.27 % + 4.4 mA		
AC VOLTAGE				A
Values can be generated for the calibration of measuring instruments	20 mV to 202 mV 10 Hz to 44 Hz 45 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 100 kHz 100 kHz to 500 kHz	920 μ V/V + 62 μ V 190 μ V/V + 62 μ V 230 μ V/V + 56 μ V 0.12 % + 84 μ V 0.46 % + 2.5 mV		
	202 mV to 2.02 V 10 Hz to 44 Hz 45 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 100 kHz	580 μ V/V + 320 μ V 180 μ V/V + 280 μ V 240 μ V/V + 450 μ V 750 μ V/V + 530 μ V		
	2.02 V to 20.2 V 10 Hz to 44 Hz 45 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 100 kHz	580 μ V/V + 3.0 mV 180 μ V/V + 2.7 mV 240 μ V/V + 4.4 mV 690 μ V/V + 5.3 mV		
	20.2 V to 202 V 30 Hz to 44 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 40 kHz	580 μ V/V + 33 mV 170 μ V/V + 28 mV 270 μ V/V + 30 mV 350 μ V/V + 53 mV		



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Generation (continued)	202 V to 1020 V 30 Hz to 44 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	640 μ V/V + 250 mV 230 μ V/V + 110 mV 290 μ V/V + 200 mV		A
For measurement of instrument outputs	10 μ V to 10 mV 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 100 kHz	230 μ V/V + 2.6 μ V 350 μ V/V + 2.6 μ V 0.12 % + 2.6 μ V		
	10 mV to 100 mV 40 Hz to 1 kHz 1 kHz to 20 kHz	82 μ V/V + 3.3 μ V 160 μ V/V + 3.3 μ V		
	100 mV to 1 V 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	82 μ V/V + 48 μ V 82 μ V/V + 27 μ V 160 μ V/V + 28 μ V 350 μ V/V + 30 μ V 920 μ V/V + 31 μ V		
	1 V to 10 V 10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	82 μ V/V + 510 μ V 82 μ V/V + 270 μ V 160 μ V/V + 270 μ V 350 μ V/V + 310 μ V 920 μ V/V + 320 μ V 0.35 % + 1.2 mV 1.2 % + 3.2 mV		
	10 V to 100 V 10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	230 μ V/V + 4.9 mV 230 μ V/V + 2.8 mV 230 μ V/V + 2.9 mV 400 μ V/V + 3.2 mV 0.14 % + 3.8 mV		
	100 V to 700 V 10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz	460 μ V/V + 52 mV 460 μ V/V + 26 mV 690 μ V/V + 28 mV		



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code	
AC CURRENT				A	
Values can be generated for the calibration of measuring Instruments	20 μ A to 202 μ A 40 Hz to 1 kHz	580 μ A/A + 180 nA			
	202 μ A to 2.02 mA 40 Hz to 1 kHz	460 μ A/A + 460 nA			
	2.02 mA 20.2 mA 40 Hz to 1 kHz	400 μ A/A + 4.6 μ A			
	20.2 mA to 202 mA 40 Hz to 1 kHz	400 μ A/A + 46 μ A			
	202 mA to 2.02 A 40 Hz to 1 kHz	460 μ A/A + 550 μ A			
	2.02 A to 20 A 40 Hz to 100 Hz	650 μ A/A + 6.0 mA			
	20 A to 30 A 40 Hz to 100 Hz	650 μ A/A + 13 mA			
	20 A to 1500 A 40 Hz to 60 Hz	0.26 % + 13 mA	Simulation using coil		
	For measurement of instrument outputs	50 nA to 100 μ A 100 Hz to 5 kHz	700 μ A/A + 46 nA		
		100 μ A to 1 mA 100 Hz to 5 kHz	350 μ A/A + 230 nA		
		1 mA to 10 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 5 kHz	0.46 % + 2.3 μ A 0.17 % + 2.3 μ A 690 μ A/A + 2.3 μ A		
		10 mA to 100 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 5 kHz	0.17 % + 24 μ A 690 μ A/A + 24 μ A 350 μ A/A + 24 μ A		
		100 mA to 1 A 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 5 kHz	0.46 % + 240 μ A 920 μ A/A + 240 μ A 0.12 % + 240 μ A		
		1 A to 3 A 10 Hz to 5 kHz	0.27 % + 580 μ A		
3 A to 5 A 10 Hz to 5 kHz		0.27 % + 11 mA			
5 A to 10 A 10 Hz to 5 kHz	0.29 % + 11 mA				



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
CAPACITANCE Values can be generated for the calibration of measuring Instruments These range values can also be measured	<i>At 1 kHz:</i> 1 nF 10 nF 20 nF 50 nF 100 nF 1 μ F 10 μ F 1 pF to 100 pF 100 pF to 1000 pF 1 nF to 10 nF 10 nF to 100 nF 0.1 μ F to 1 μ F 1 μ F to 10 μ F	3.5 pF 31 pF 64 pF 150 pF 290 pF 4.6 nF 69 nF 0.12 % + 0.12 pF 0.12 % + 0.33 pF 0.12 % + 3.1 pF 0.12 % + 31 pF 0.12 % + 310 pF 0.12 % + 0.31 nF	Comparison against LCR bridge	A
INDUCTANCE Values can be generated for the calibration of measuring instruments These range values can also be measured	<i>At 1 kHz:</i> 1 mH 10 mH 100 mH 1 H 0.1 mH to 1 mH 1 mH to 10 mH 10 mH to 100 mH 0.1 H to 1 H	5.9 μ H 58 μ H 580 μ H 5.8 mH 0.12 % + 0.39 μ H 0.27 % + 3.9 μ H 0.12 % + 35 μ H 0.12 % + 300 μ H	Comparison against LCR bridge	A
FREQUENCY Value can be generated for the calibration of measuring instruments For generating a stimulus that can be applied to measuring instruments also for measuring a stimulus provided by the device being calibrated Rotational speed - Optical Measurement Generation	10 MHz reference 1 Hz to 30 MHz 30 MHz to 4 GHz 10 RPM to 99.99 RPM 100 RPM to 999.9 RPM 1000 RPM to 99999 RPM 60 RPM to 3000 RPM 3000 RPM to 60000 RPM	1.0 part in to 10^{12} 1.5 parts in to $10^{12} + 0.60 \mu$ Hz 2.0 parts in to 10^{12} 2.3 RPM 2.4 RPM 3.3 RPM 0.12 RPM 1.2 RPM	Frequency may also be expressed as time; 1/f for repetitive signals, in terms of seconds or other units such as RPM.	A



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE SIMULATION				A
PT 100	-200 °C to +800 °C	0.065 °C		
Ambient	17 °C to 23 °C	0.20 °C	Calibration of Cold Junction	
Reference (CJC) juncton compensation INCLUDED				
Base Thermocouples				
Type E	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1000 °C	0.22 °C 0.21 °C 0.22 °C		
Type J	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1200 °C	0.24 °C 0.21 °C 0.22 °C		
Type K	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1372 °C	0.27 °C 0.22 °C 0.24 °C		
Type N	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1300 °C	0.35 °C 0.24 °C 0.24 °C		
Type T	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 400 °C	0.27 °C 0.23 °C 0.21 °C		
Noble thermocouples				
Type B	600 °C to 1820 °C	0.52 °C		
Type R	-50 °C to 0 °C 0 °C to 400 °C 400 °C to 1767 °C	0.78 °C 0.57 °C 0.36 °C		
Type S	-50 °C to 0 °C 0 °C to 400 °C 400 °C to 1767 °C	0.65 °C 0.55 °C 0.40 °C		



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Temperature indicators and calibrators by electrical simulation Reference (CJC) junction compensation EXCLUDED				A
Base Thermocouples				
Type E	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1000 °C	0.22 °C 0.19 °C 0.20 °C		
Type J	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1200 °C	0.24 °C 0.19 °C 0.20 °C		
Type K	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1372 °C	0.26 °C 0.21 °C 0.22 °C		
Type N	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1300 °C	0.35 °C 0.22 °C 0.22 °C		
Type T	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 400 °C	0.26 °C 0.21 °C 0.20 °C		
Noble thermocouples				
Type B	600 °C to 1820 °C	0.52 °C		
Type R	-50 °C to 0 °C 0 °C to 400 °C 400 °C to 1767 °C	0.77 °C 0.56 °C 0.35 °C		
Type S	-50 °C to 0 °C 0 °C to 400 °C 400 °C to 1767 °C	0.65 °C 0.55 °C 0.39 °C		



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
EQUIPMENT FOR IEE 17 TH / 18 TH EDITION WIRING TESTING				
LOOP TESTERS	Nominal applied resistances			A
AC Resistance at 50 Hz	0.05 Ω	4.7 m Ω		
	0.10 Ω	4.8 m Ω		
	0.21 Ω	4.9 m Ω		
	0.32 Ω	5.1 m Ω		
	0.5 Ω	5.6 m Ω		
	1 Ω	8.6 m Ω		
	5 Ω	31 m Ω		
	10 Ω	59 m Ω		
	100 Ω	580 m Ω		
	1 k Ω	5.9 Ω		
CONTINUITY TESTERS				A
DC Resistance	20 m Ω	29 m Ω		
	200 m Ω to 2 Ω	29 m Ω		
	4 Ω	31 m Ω		
	6 Ω	34 m Ω		
	8 Ω	37 m Ω		
	10 Ω	41 m Ω		
	20 Ω	65 m Ω		
	100 Ω	290 m Ω		
	1 k Ω	2.9 Ω		
Continuity Current Measurement	10 mA	1.1 mA		
	100 mA	1.7 mA		
	200 mA	3.1 mA		
	300 mA	4.6 mA		



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
INSULATION TESTERS				A
DC Resistance	10 k Ω 20 k Ω 30 k Ω 40 k Ω 60 k Ω 100 k Ω 200 k Ω 400 k Ω 600 k Ω 1 M Ω 2 M Ω 3 M Ω 4 M Ω 5 M Ω 6 M Ω 7 M Ω 8 M Ω 9 M Ω 10 M Ω 20 M Ω 30 M Ω 40 M Ω 50 M Ω 60 M Ω 70 M Ω 80 M Ω 90 M Ω 100 M Ω 200 M Ω 400 M Ω 600 M Ω 800 M Ω 1 G Ω 10 G Ω	12 Ω 23 Ω 35 Ω 46 Ω 69 Ω 120 Ω 230 Ω 460 Ω 690 Ω 1.2 k Ω 2.3 k Ω 3.5 k Ω 4.6 k Ω 58 k Ω 69 k Ω 81 k Ω 92 k Ω 100 k Ω 120 k Ω 230 k Ω 350 k Ω 460 k Ω 580 k Ω 690 k Ω 810 k Ω 930 k Ω 1.0 M Ω 1.2 M Ω 2.8 M Ω 5.6 M Ω 8.5 M Ω 11 M Ω 14 M Ω 580 M Ω		
DC Voltage	50 V 100 V 150 V 200 V 250 V 500 V 1000 V	1.1 V 1.5 V 2.0 V 2.5 V 3.0 V 5.9 V 12 V		



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
EARTH BOND TESTERS				A
AC Resistance at 50 Hz	Nominal applied resistance			
	0.04 Ω	4.7 m Ω		
	0.1 Ω	4.8 m Ω		
	0.15 Ω	4.8 m Ω		
	0.27 Ω	5.0 m Ω		
	0.38 Ω	5.2 m Ω		
	0.55 Ω	5.8 m Ω		
	1 Ω	7.8 m Ω		
	5 Ω	30 m Ω		
	10 Ω	59 m Ω		
	100 Ω	580 m Ω		
	1 k Ω	5.8 Ω		
AC Current at 50 Hz				
	100 mA	7.3 mA		
	200 mA	7.9 mA		
	400 mA	9.9 mA		
	4 A	100 mA		
	8 A	160 mA		
	10 A	190 mA		
	20 A	440 mA		
LEAKAGE TESTERS				A
DC Current				
	2 mA	36 μ A		
	5 mA	82 μ A		
	10 mA	130 μ A		
RCD TESTERS				A
RCD Trip Time				
	20 ms	680 μ s		
	40 ms	680 μ s		
	100 ms	680 μ s		
	200 ms	680 μ s		
	390 ms	680 μ s		
	900 ms	8.1 ms		
RCD Trip Current at 50 Hz				
	10 mA	620 μ A		
	30 mA	1.7 mA		
	90 mA	5.2 mA		
	100 mA	5.8 mA		
	110 mA	6.4 mA		
	150 mA	17 mA		
	300 mA	17 mA		
	1 A	58 mA		
	2 A	120 mA		
AC Voltage Source at 50 Hz				
	100 V	0.37 V		
	200 V	0.45 V		
	230 V	0.65 V		
	300 V	0.82 V		
	400 V	0.99 V		
Line Voltage Measurement	200 V to 260 V	2.4 V		



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
PRESSURE				A
Gas Pressure (Gauge)			Methods consistent with EURAMET CG17. Calibration of devices with an electrical output may be undertaken.	
Calibration of pressure indicating instruments and gauges	-95 kPa to -3.5 kPa 3.5 kPa to 100 kPa 100 kPa to 2.5 MPa 2.5 MPa to 12 MPa	0.0073 % 0.0044 % 0.0044 % 0.0057 %	Calibration using deadweight testers. Absolute pressures can be generated over these ranges attracting an additional uncertainty of 7.5 Pa.	
Calibration of pressure indicating instruments and gauges	-2.4 kPa to +2.4 kPa 3.5 kPa to 100 kPa 100 kPa to 2.1 MPa -95 kPa to 0 Pa 0 Pa to 21 MPa	0.078 % + 1.3 Pa 0.004 6 % + 13 Pa 0.002 5 % + 90 Pa 0.018% + 1.9 kPa 0.014 % + 1.9 kPa	Calibration using pressure controllers.	
Gas Pressure (Absolute)				
Calibration of pressure indicating instruments and gauges	75 kPa to 115 kPa 3.5 kPa to 100 kPa 3.5 kPa to 200 kPa 3.5 kPa to 800 kPa 3.5 kPa to 2.1 MPa 3.5 kPa to 7 MPa 100 kPa to 41.4 MPa	7.5 Pa 0.0034 % + 9.5 Pa 0.0034 % + 10 Pa 0.0031 % + 140 Pa 0.0030 % + 77 Pa 0.0030 % + 840 Pa 0.0041 % + 4.8 kPa	Calibration using pressure controllers.	
Hydraulic Pressure (Gauge)				
Calibration of pressure indicating instruments and gauges	0.6 MPa to 6.0 MPa 6 MPa to 70 MPa 70 MPa to 138 MPa 138 MPa to 344 MPa	0.0062 % 0.0075 % 0.0073 % 0.0086 %	Calibration using deadweight testers. Absolute pressures can be generated over these ranges attracting an additional uncertainty of 7.5 Pa.	
TEMPERATURE				A
PRTs and Sensors with indicators	-95 °C to 140 °C 140 °C to 660 °C 140 °C to 660 °C	0.056 °C 0.15 °C 0.10 °C	Calibrations performed in a Metal block	
	-80 °C to 100 °C 100 °C to 150 °C 150 °C to 250 °C	0.019 °C 0.031 °C 0.040 °C	Calibrations performed in liquid bath	
	0 °C	0.010 °C	Ice point	
	0.01 °C	0.0050 °C	Triple point of water	
Metal block calibrators and portable liquid baths	-95 °C to +660 °C	Uncertainty as for sensor and indicator		



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
HUMIDITY				A
Dew point	-20 °C to 60 °C	0.20 °C	By comparison with a chilled mirror hygrometer	
Relative humidity	Example conditions	Corresponding to dew-point and temperature range and uncertainties	By comparison with a chilled mirror hygrometer and PRTs	
	At 0 °C 5 %rh 50 %rh 90 %rh	0.16 %rh 0.85 %rh 1.5 %rh		
	At 23 °C 5 %rh to 10 %rh 10 %rh to 50 %rh 50 %rh to 95 %rh	0.15 %rh 0.72 %rh 1.2 %rh		
Relative humidity	At 60 °C 5 %rh 50 %rh 90 %rh	0.14 %rh 0.61 %rh 1.0 %rh		
TEMPERATURE IN AIR	0 °C to 60 °C	0.15 °C		A



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
DIMENSIONAL CALIBRATIONS	RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED			
LENGTH			NOTES 1 All linear calibrations may be given in inch units.	
Plain Plug Gauges (Parallel)	1 to 50 diameter 50 to 100 100 to 200 200 to 300	0.80 1.5 2.0 3.0	on diameter.	B
Length Gauges, Flat and Spherical Ended (excluding Length Bars)	25 to 1000	$1.0 + (8.0 \times \text{length in m})$	Comparison to gauge blocks using a length measuring machine.	B
ANGLE			2 The uncertainty quoted is for the departure from flatness, straightness, parallelism or squareness, i.e., the distance separating the two parallel planes which just enclose the surface under consideration.	
Squares Blade Type	50 to 300 300 to 450	3.0 on squareness 5.0 See Note 2	BS 939:2007 Comparison to master square.	B
MEASURING INSTRUMENTS AND MACHINES				
Micrometers			Comparison to length standards	
External	0 to 1000	Heads 2.0 between any two points	BS 870:2008	B
Internal Micrometers	0 to 900	Setting and extension rods	BS 959:2008	B
Depth Micrometers	0 to 300	$1.0 + (8.0 \times \text{length in m})$	BS 6468:2008	B
Vernier, dial and digital type gauges			Comparison to length standards.	B
Calliper	0 to 1000	Overall performance $10 + (30 \times \text{length in m})$	As BS 887:2008	
Height	0 to 1000	Overall performance $10 + (10 \times \text{length in m})$	ISO13225:2012 and BS 1643:2008	
Depth	0 to 600	Overall performance $10 + (30 \times \text{length in m})$	As BS 6365:2008	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
MEASURING INSTRUMENTS AND MACHINES (continued)				
Dial Gauges and Dial Test Indicators	0 to 50	1.0	BS 907:2008 and BS 2795:1981 using a length measuring machine.	B
Surface Plates Granite Cast Iron	160 x 100 to 4000 x 4000 Flatness of working surface: Local variation of working surface:	1.5 + (0.80 x diagonal in m) See Note 2 2.7	BS 817:2008 and above using an electronic level and variation gauge.	B and C
Feeler Gauges	0.025 to 1.0	2.0	BS 957:2008 using a length measuring machine.	B
ELECTRICAL CALIBRATION				
DC RESISTANCE Specific values				B
Measurement	10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω 1 G Ω 10 G Ω 100 G Ω 1 T Ω	14 $\mu\Omega/\Omega$ 9.6 $\mu\Omega/\Omega$ 8.0 $\mu\Omega/\Omega$ 8.8 $\mu\Omega/\Omega$ 9.6 $\mu\Omega/\Omega$ 24 $\mu\Omega/\Omega$ 110 $\mu\Omega/\Omega$ 440 $\mu\Omega/\Omega$ 0.40 % 0.59 % 2.0 % 1.4 %	Using digital multimeter.	
Other values Measurement	0 Ω to 1 Ω 1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	66 $\mu\Omega/\Omega$ + 2.0 $\mu\Omega$ 73 $\mu\Omega/\Omega$ + 2.0 $\mu\Omega$ 200 $\mu\Omega/\Omega$ + 2.0 $\mu\Omega$ 13 $\mu\Omega/\Omega$ + 2.0 $\mu\Omega$ 14 $\mu\Omega/\Omega$ 76 $\mu\Omega/\Omega$ 37 $\mu\Omega/\Omega$ 150 $\mu\Omega/\Omega$ 0.13 % 0.14 %	Using digital multimeter.	B



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
DC RESISTANCE (continued)				B
Generation				
Specific values	10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω	6.2 m Ω 8.0 m Ω 28 m Ω 160 m Ω 2.9 Ω 45 Ω 1.3 k Ω 260 k Ω	Using multifunction calibrator or decade resistance box.	
DC VOLTAGE				B
Measurement				
Specific values	100 mV 1 V 10 V 100 V 1000 V	11 μ V/V 9.4 μ V/V 9.4 μ V/V 12 μ V/V 12 μ V/V	Using digital multimeter.	
Other values	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	72 μ V/V + 2.0 μ V 10 μ V/V + 2.0 μ V 9.5 μ V/V + 2.0 μ V 19 μ V/V 12 μ V/V	Using digital multimeter.	
Generation	1 kV to 20 kV 20 kV to 30 kV	0.47 kV 0.35 %	Using high voltage divider.	
Generation	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V	18 μ V/V + 3.0 μ V 11 μ V/V + 5.0 μ V 10 μ V/V + 50 μ V 14 μ V/V + 530 μ V	Using multifunction calibrator.	
DC CURRENT	200 V to 1025 V	14 μ V/V + 4.0 mV		
Measurement				B
Specific values	1 μ A 10 μ A 100 μ A 1 mA 10 mA 100 mA 1 A	45 μ A/A 25 μ A/A 24 μ A/A 24 μ A/A 24 μ A/A 41 μ A/A 87 μ A/A	Using digital multimeter.	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
DC CURRENT (continued) Measurement (continued) Other values	0 μ A to 1 μ A 1 μ A to 10 μ A 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A	470 pA 120 μ A/A 96 μ A/A 63 μ A/A 64 μ A/A 71 μ A/A 180 μ A/A	Using digital multimeter.	B
Generation	0 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A 20 A to 30 A	0.020 % + 12 nA 0.010 % + 1.2 μ A 0.010 % + 12 μ A 0.040 % + 3.3 μ A 0.020 % + 120 μ A 0.040 % + 2.0 mA 0.060 % + 2.0 mA	Using multifunction calibrator.	B
AC VOLTAGE Measurement Specific values	30 to 1000 A	0.75 % + 1.4 mA	For the calibration of clamp on ammeters and similar devices, using multi-turn method.	B
Other values	<i>At 1 kHz</i> 10 mV 100 mV <i>40 Hz to 1 kHz</i> 1 V 10 V 100 V 700 V <i>1 kHz to 100 kHz</i> 1 V 10 V 100 V	190 μ V/V 73 μ V/V 64 μ V/V 65 μ V/V 160 μ V/V 310 μ V/V 620 μ V/V 620 μ V/V 930 μ V/V	Using digital multimeter.	B
	<i>At 1 kHz</i> 1 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V <i>40 Hz to 1 kHz</i> 1 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 700 V <i>1 kHz to 100 kHz</i> 1 V to 10 V 10 V to 100 V	0.090 % 0.020 % 0.020 % 0.13 % 0.025 % 0.025 % 0.025 % 0.033 % 0.052 % 0.095 % 0.14 %		



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
AC VOLTAGE (continued) Measurement (continued)	<i>At 50 Hz</i> 700 V to 20 kV 20 kV to 28 kV	1.2 kV 1.6 kV	Using high voltage divider.	B
Generation	<i>45 Hz to 20 kHz</i> 10 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1020 V	0.023 % + 20 μ V 0.024 % + 260 μ V 0.024 % + 2.2 mV 0.035 % + 37 mV 0.35 % + 250 mV	Using multifunction calibrator.	
AC CURRENT Measurement Specific Values	<i>At 1 kHz</i> 100 μ A 1 mA <i>45 Hz to 1 kHz</i> 10 mA 100 mA 1 A	0.056 % 0.051 % 0.050 % 0.050 % 0.080 %	Using digital multimeter.	B
Other values	<i>At 1 kHz</i> 5 μ A to 100 μ A 100 μ A to 1 mA <i>45 Hz to 1 kHz</i> 5 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A	0.24 % 0.16 % 0.36 % 0.25 % 0.24 % 0.24 % 0.26 %	Using digital multimeter.	
Generation	<i>45 Hz to 1 kHz</i> 20 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 30 A 30 A to 1000 A	0.090 % + 180 nA 0.070 % + 490 μ A 0.50 % + 4.0 μ A 0.50 % + 36 μ A 0.070 % + 340 μ A 0.35 % + 6.8 mA 0.83 % + 6.8 mA	Using multifunction calibrator. For the calibration of clamp on ammeters and similar devices, using multi-turn method.	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
FREQUENCY	1 Hz to 1 MHz 1 MHz to 2.1 GHz	21 in $10^8 + 2.0$ mHz 21 in 10^8	Using frequency counter.	B
Tachometer calibration	10 rpm to 50000 rpm	1.2 rpm	Using optical technique.	
Elapsed time	0 ms to 390 ms 391 ms to 100 s	1.0 ms 8.0 ms	Using counter timer.	B
17 TH EDITION TYPE EQUIPMENT Earth Loop	0.05 Ω 0.1 Ω 0.22 Ω 0.33 Ω 0.5 Ω 1 Ω 5 Ω 10 Ω 100 Ω 1 k Ω	10 m Ω 11 m Ω 8.0 m Ω 8.0 m Ω 8.0 m Ω 10 m Ω 30 m Ω 59 m Ω 580 m Ω 5.8 m Ω	Using dedicated calibrator.	B
RCD testers				B
Trip current	At 50 Hz 3 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 3 A	620 μ A 5.8 mA 59 mA 120 mA	Up to 5 seconds.	
Trip Time	At 30 mA, 50 Hz 10 ms to 390 ms 390 ms to 1 s	1.0 ms 8.1 ms		
Earth leakage current	0.2 mA to 7.7 mA	15 μ A		B
PAT Testers Earth bond current	At 50 Hz 100 mA 100 mA to 10 A 10 A to 30 A	8.0 mA 190 mA 520 mA		B
Earth Bond resistance Nominal values	0.05 Ω 0.1 Ω 0.22 Ω 0.33 Ω 0.5 Ω 1 Ω 5 Ω 10 Ω 100 Ω 1 k Ω	7.5 m Ω 7.5 m Ω 7.6 m Ω 7.7 m Ω 8.0 m Ω 9.5 m Ω 30 m Ω 58 m Ω 580 m Ω 5.8 Ω		B



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
17 TH EDITION TYPE EQUIPMENT (continued)				B
Insulation resistance Nominal source values	100 kΩ 500 kΩ 1 MΩ 5 MΩ 10 MΩ 20 MΩ 50 MΩ 100 MΩ	12 kΩ 12 kΩ 12 kΩ 21 kΩ 37 kΩ 72 kΩ 180 kΩ 350 kΩ		B
Load Tests	3 kVA	2.5 %		B
Flash tests	At 50 Hz 700 V to 1.9 kV	1.5 % + 5.0 V		
ELECTRICAL SIMULATION OF TEMPERATURE				B
Ambient temperature CJC source CJC measurement	17 °C to 23 °C	0.11 °C 0.32 °C	In support of cold junction measurements.	
Temperature simulators and indicators, calibration by electrical simulation				
Base metal thermocouples				
Type K	-200 °C to -100 °C -100 °C to +1300 °C	0.32 °C 0.30 °C	Excluding cold junction compensation.	
	-200 °C to -100 °C -100 °C to +1300 °C	0.39 °C 0.38 °C	Including cold junction compensation.	
Type J	-210 °C to -100 °C -100 °C to +1200 °C	0.32 °C 0.32 °C	Excluding cold junction compensation.	
	-210 °C to -100 °C -100 °C to +1200 °C	0.37 °C 0.39 °C	Including cold junction compensation.	
Type N	-200 °C to -100 °C -100 °C to +1300 °C	0.49 °C 0.29 °C	Excluding cold junction compensation.	
	-200 °C to -100 °C -100 °C to +1300 °C	0.54 °C 0.37 °C	Including cold junction compensation.	
Type T	-250 °C to -150 °C -150 °C to +400 °C	0.71 °C 0.14 °C	Excluding cold junction compensation.	
	-250 °C to -150 °C -150 °C to +400 °C	0.75 °C 0.27 °C	Including cold junction compensation.	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Noble metal thermocouples Type B Type R, Type S	-			
	600 °C to 800 °C	0.85 °C	Excluding cold junction compensation.	
	800 °C to 1820 °C	0.76 °C		
	600 °C to 800 °C	0.88 °C	Including cold junction compensation.	
	800 °C to 1820 °C	0.79 °C		
	0 °C to 250 °C	0.93 °C	Excluding cold junction compensation.	
	250 °C to 1760 °C	0.60 °C		
	0 °C to 250 °C	0.96 °C	Including cold junction compensation.	
250 °C to 1760 °C	0.64 °C			
END				



8239

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 015 Issue date: 18 December 2024

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$